

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	103083
Date Issue	JUL2026
Expiration Date	AUG2027

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)	
Mouse	☒ Yes	☐ No
Rat	☒ Yes	☐ No
Hamster	☒ Yes	☐ No
Rabbit	☒ Yes	☐ No
Human	☒ Yes	☐ No

Section B: MFI*

Population	MFI Specification	Results (Pass/Fail)	
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone: SK1)	≥400,000	☒ Pass	☐ Fail
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC; clone: PC61)	≥400,000	☒ Pass	☐ Fail
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone: 145-2C11)	≥400,000	☒ Pass	☐ Fail
Rabbit Positive (stained using Rabbit anti-human kappa FITC)	≥400,000	☒ Pass	☐ Fail
Human Positive (stained with Human IgG1 FITC; clone: QA16A12)	≥400,000	☒ Pass	☐ Fail

SpectraComp Flex XT (SSB-22-A) Certificate of Analysis

Slingshot Biosciences

COA19-3

SpectraComp Flex XT (SSB-22-A) Lot 103083 Certificate of Analysis

Version: 1.0 | Effective: Jul 20 2026

***when analyzed using Cytex Aurora 5 Laser Cytometer under recommended gain settings**

Signed by:

Camille Shah

D62A2FF04F7F446...

Approved by: _____

Quality Assurance