

Scope The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.

Product LOT 211003

Date Issued: 07NOV2021

Section A: Reactivity

Species	Reactivity Confirmed
Mouse	Yes
Rat	Yes
Hamster	Yes

Section B: MFI*

Population	MFI specification	Results
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	> 5e5	PASS
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	> 5e5	PASS
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	> 5e5	PASS

*when analyzed using on Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by:



Tam Truong, Vice President Slingshot Dx



SLINGSHOT

TITLE	SSB-05-A/B SpectraComp CoA Lot 211003
FILE NAME	Lot 211003- SSB-0...ectraComp COA.pdf
DOCUMENT ID	11ae26110ad7f1b6dcb93ffa6ca42ebfca0929bc
AUDIT TRAIL DATE FORMAT	MM / DD / YYYY
STATUS	● Signed

Document History



SENT

11 / 01 / 2022

23:15:44 UTC

Sent for signature to Tam Truong (tam@slingshotbio.com) from
 brittney@slingshotbio.com
 IP: 76.77.180.58



VIEWED

11 / 01 / 2022

23:16:05 UTC

Viewed by Tam Truong (tam@slingshotbio.com)
 IP: 47.156.225.5



SIGNED

11 / 01 / 2022

23:16:19 UTC

Signed by Tam Truong (tam@slingshotbio.com)
 IP: 47.156.225.5



COMPLETED

11 / 01 / 2022

23:16:19 UTC

The document has been completed.