

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	103141
Date Issued	AUG2026
Expiration Date*	AUG2028

***Product has a shelf life of six months from the date of thaw. Follow the Expiration Date if it occurs before shelf-life expiration.**

Section A: Staining Dyes


Stain	Staining Confirmed (Yes/No)	
7AAD	☒ Yes	☐ No
Ghost	☒ Yes	☐ No

Section B: MFI**

Population	MFI Specification	Results (Pass/Fail)	
Positive for 7AAD	≥20,000	☒ Pass	☐ Fail
Positive for Ghost	≥200,000	☒ Pass	☐ Fail

****when analyzed using Cytek Aurora 5 Laser Cytometer under recommended gain settings**

Approved by: _____

Signed by:

D62A2FF04F7F446...

Quality Assurance