

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for <i>In Vitro</i> Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	103211
Date Issued	SEP2026
Expiration Date	FEB2028

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)	
Mouse	☒ Yes	☐ No
Rat	☒ Yes	☐ No
Hamster	☒ Yes	☐ No

Section B: MFI*

Population	MFI Specification	Pass/Fail	
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	≥400,000	☒ Pass	☐ Fail
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	≥400,000	☒ Pass	☐ Fail
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	≥400,000	☒ Pass	☐ Fail

* when analyzed using Cytex Aurora 5 Laser Cytometer under recommended gain settings

Signed by:

Camille Shah

D62A2FF04F7F446...

Approved by: _____

Quality Assurance