1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for *In Vitro* Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	103014
Date Issued	SEP2026
Expiration Date	FEB2028

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)	
Mouse	☑ Yes	☐ No
Rat	☑ Yes	☐ No
Hamster	☑ Yes	☐ No

Section B: MFI*

Population	MFI Specification	Pass/Fail	
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	≥400,000	☑ Pass	☐ Fail
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC; clone PC61.5)	≥400,000	☑ Pass	☐ Fail
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	≥400,000	☑ Pass	☐ Fail

*** when analyzed using Cytek Aurora 5 Laser Cytometer under recommended gain settings**

Approved by: Signed by: Camille Shah D62A2FF04F7F446...

Quality Assurance

Property of Slingshot Biosciences. **CONFIDENTIAL**. Unauthorized distribution or copying prohibited. Generated by Camille Shah on 05 Oct 2026 at 11:27:13 (UTC -0700) Controlled copy.