

# 1. Certificate of Analysis

<b>Scope</b>	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
<b>Product Lot</b>	100439
<b>Date Issued</b>	Feb 2023
<b>Expiration Date</b>	Feb 2024

## Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)
Mouse	Yes
Rat	Yes
Hamster	Yes

## Section B: MFI\*

Population	MFI specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti-human CD8 FITC; clone-SK1)	> 500,000	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5)	> 500,000	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	> 500,000	Pass

\*when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings

Approved by: Jonathan Yang 02 / 15 / 2023  
Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: 11-0530v2