Version: 2.0

COA-7

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	100298
Date Issued	06Feb2023
Expiration Date	06Feb2024

Section A: Reactivity

Species	Reactivity Confirmed (Yes/No)	
Mouse	Yes	
Rat	Yes	
Hamster	Yes	

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Mouse Positive (stained using Mouse IgG1 Anti- human CD8 FITC; clone-SK1)	> 500,000	Pass
Rat Positive (stained using Rat IgG1 Anti-mouse CD25 FITC clone PC61.5	> 500,000	Pass
Hamster Positive (stained using Armenian Hamster Anti-mouse CD3e FITC; clone 145-2C11)	> 500,000	Pass

*when analyzed using Cytek Aurora 5 laser Cytometer under recommended gain settings

Approved by: _	Jonathan Yang	02 / 08 / 2023	

Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: 11-0530v2



Title 100298

File name COA-7-v2.0.pdf

Document ID 59288e1143d91671afb22a266e72840354d4bfc7

Audit trail date format MM / DD / YYYY

Status • Signed

Document History

O2 / 08 / 2023 Sent for signature to Jonathan Yang

SENT 21:43:56 UTC (john.yang@slingshotbio.com) from john.yang@slingshotbio.com

IP: 76.77.180.58

O2 / 08 / 2023 Viewed by Jonathan Yang (john.yang@slingshotbio.com)

VIEWED 21:43:57 UTC IP: 76.77.180.58

SIGNED 21:44:10 UTC IP: 76.77.180.58

7 02 / 08 / 2023 The document has been completed.

COMPLETED 21:44:10 UTC