

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product Lot	100793
Date Issued	Jun2023

Population	MFI specification	Results (Pass/Fail)
SpectraComp eGFP Negative	<3000	Pass
SpectraComp eGFP Low	10k - 30k	Pass
SpectraComp eGFP Medium	100k - 250k	Pass
SpectraComp eGFP High	1M - 2.5M	Pass

***when analyzed using Cytex Aurora 5 laser Cytometer under recommended gain settings**

Approved by: Jonathan Yang 06 / 13 / 2023
Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: 11-0630v1