

1. Certificate of Analysis

Scope	The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.
Product LOT	100697
Date Issued	May2023
Expiration Date	May2024

Section A: Staining Dyes

Stain	Staining Confirmed (Yes/No)
7AAD	Yes
Ghost	Yes

Section B: MFI*

Population	MFI specification	Results (Pass/Fail)
Negative for 7AAD	≤ 500	Pass
Positive for 7AAD	$\geq 20,000$	Pass
Negative for Ghost	$\leq 15,000$	Pass
Positive for Ghost	$\geq 200,000$	Pass

*when analyzed using on Cytek Aurora 5 laser Cytometer under recommended gain settings

Approved by: Jonathan Yang 05 / 03 / 2023
Jonathan Yang / Director, Quality Assurance and Regulatory Affairs

Slingshot Biosciences Internal Part Number: SSB-11-0552