

Scope The product details below have not been cleared by the FDA for In Vitro Diagnostic use. These values are provided for Research Use Only. Not for use in diagnostic procedures.

Product LOT 100218

Date Issued: Sep2022

Section A: Staining Dyes

Stain	Staining Confirmed
7AAD	Yes
Ghost	Yes

Section B: MFI*

Population	MFI specification	Results
Negative for 7AAD	<100	Pass
Positive for 7AAD	>20,000	Pass
Negative for Ghost	<15,000	Pass
Positive for Ghost	>200,000	Pass

*when analyzed using on Cytek Aurora 5 laser Cytometer under recommended gain settings

Approved by:

Jonathan Yang

Jonathan Yang / Director, Quality Assurance and Regulatory Affairs



SLINGSHOT